

ALCOTEST CERTIFICATIONS INDEX # 10

06-02-2026

ALCOTEST 9510

SERIAL # ARMJ-0142

- ITEM # 1 9510 PARAMETER REPORT
- ITEM # 2 9510 CERT REPORT – WET ADJUST (PART I)
- ITEM # 3 9510 CERT REPORT – DRY ADJUST (PART II)
- ITEM # 4 9510 CERT REPORT – LINEARITY (PART III)
- ITEM # 5 9510 CYLINDER INSTALLATION REPORT – INLET 1
- ITEM # 6 9510 CYLINDER INSTALLATION REPORT – INLET 2
- ITEM # 7 CERT OF ANALYSIS, CALGAS LOT# 302-403119242
- ITEM # 8 CERT OF ANALYSIS, CALGAS LOT# 302-403386733
- ITEM # 9 CERT OF ACCURACY 9510, SERIAL # ARMJ-0142
- ITEM # 10 CERT OF ANALYSIS SIMULATOR SOLUTION LOT # 25190
- ITEM # 11 CERT OF CALIBRATION, CERT/SO#5-G0P3R-40-1 Revision 0 (2 pgs)
- ITEM # 12 CERT OF CALIBRATION, CERT/SO#5-F8B2G-40-1 Revision 0 (6 pgs)
- ITEM # 13 CERT OF ANALYSIS, CALGAS LOT# 302-403034217
- ITEM # 14 CERT OF ANALYSIS, CALGAS LOT# 302-402999655
- ITEM # 15 CERT OF ANALYSIS, CALGAS LOT# 403008098
- ITEM # 16 CERT OF ANALYSIS, CALGAS LOT# 302-402997176
- ITEM # 17 CERT OF ANALYSIS, CALGAS LOT# 302-403001320
- ITEM # 18 NJSP CERTIFICATION OFFICER CREDENTIALS

ALCOTEST 9510 PARAMETER REPORT

Equipment

Serial No.: ARMJ-0142
Firmware: 8326739 1.5
WinCE application: 8326738 2.9
Configuration: 8326737 3.10

Date: 06/02/2026
Time: 06:23:06

Parameter

min. blow time	5.0	s
min. breath volume for females of age 60+	1.2	L
min. breath volume for all other	1.5	L
min. blow flow	4.5	L/min
plateau detection limit	4	%
plateau detection start conc.	70	microgram/L
neg. flow detection (part. vacuum)	10.0	hPa
neg. flow detection sensitivity	10	
cal. gas abort volume	0.4	L
result-to-zero limit	0.0050	%BAC
ambient air check limit	0.0049	%BAC
interference det. d-criterion limit abs.	38	microgram/L
interference det. d-criterion limit rel.	10.0	%
interference det. t-criterion limit abs.	8	microgram/L
interference det. t-criterion limit rel.	2.1	%
IR CO2 offset	10	microgram/L
IR H2O offset	4	microgram/L
EC H2O offset	0	microgram/L
Value-based EC aging comp. on/off (1/0)	0	
Time-based EC aging comp. on/off (1/0)	0	
Time-based EC aging comp. per month	0.2	%
Time-based EC aging comp. maximum	3.0	%
EC fatigue comp. max. sum	15000	
EC fatigue comp. factor	50	
EC fatigue comp. minutes	180	
mouth alc. mark limit	500	
mouth alc. lower limit	30	
mouth alc. slope	6	
mouth alc. zero limit	50	
mouth alc. max. neg. sum	6	
mouth alc. max. 2nd derivative	35	

ALCOTEST 9510 CERTIFICATION REPORT - WET ADJUST (PART I)
Toms River Twp

Equipment

Inst. Model No.: ALCOTEST 9510 Serial No.: ARMJ-0142
Firmware: 8326739 1.5 Config.: 8326737 3.10 WinCE: 8326738 2.9

Wet Adjust Record

Wet Adjust File No.: 24 Wet Adjust Date: 06/02/2026 Wet Adjust No.: 1
Wet Adjust Time: 07:08:02

Concentration: 0.100 %
Adjusting Unit: X-Cal 2000 Adj. Unit Ser. No.: ARRL-0020 Adj. Unit Exp.: 01/15/2027
Solution Lot No.: 25190 Soln. Bottle No.: 573 Adjust Soln. Exp.: 06/10/2027

Preadjust Simulator Temp.: 34.01 degree C
Postadjust Simulator Temp.: 34.00 degree C

Result

Procedure completed successfully.

Coordinator

Last Name: Bellay - First Name: David MI: M. Badge No.: 8112

On this date, I certified the above instrument in accordance with the Alcotest 9510 operator training and procedures established by the NJSP Office of Forensic Sciences.



Signed:

Date: 06/02/2026

ID: 50

ALCOTEST 9510 CERTIFICATION REPORT - DRY ADJUST (PART II)
Toms River Twp

Equipment

Inst. Model No.: ALCOTEST 9510 Serial No.: ARMJ-0142
Firmware: 8326739 1.5 Config.: 8326737 3.10 WinCE: 8326738 2.9

Dry Adjust Record

Dry Adjust File No.: 25 Dry Adjust Date: 06/02/2026 Dry Adjust No.: 1
Dry Adjust Time: 07:26:06

Concentration: 0.100 %
Dry Gas Lot No.: 302-403034217 Adjust Gas Exp.: 04/30/2027
Barom. Model No.: Mensor CPG2300 Barom. Serial No.: 41001275 Barom. Cert. Exp.:08/23/2026
Pre-adjust Amb. Pressure: 1017 hPa Post-adjust Amb. Pressure: 1020 hPa

Result

Procedure completed successfully.

Coordinator

Last Name: Bellay - First Name: David MI: M. Badge No.: 8112

On this date, I certified the above instrument in accordance with the Alcotest 9510 operator training and procedures established by the NJSP Office of Forensic Sciences.



Signed:

Date: 06/02/2026

ID: 50

ALCOTEST 9510 CERTIFICATION REPORT - LINEARITY (PART III)
Toms River Twp

Equipment

Inst. Model No.: ALCOTEST 9510 Serial No.: ARMJ-0142
Firmware: 8326739 1.5 Config.: 8326737 3.10 WinCE: 8326738 2.9

Linearity Record

Linearity File No.: 26 Lin. Date: 06/02/2026 Lin. No.: 1

0.040% Dry Gas Lot No.: 302-402999655 Adjust. Gas Exp.: 03/20/2027
0.080% Dry Gas Lot No.: 302-403008098 Adjust. Gas Exp.: 04/01/2027
0.160% Dry Gas Lot No.: 302-402997176 Adjust. Gas Exp.: 03/19/2027
0.300% Dry Gas Lot No.: 302-403001320 Adjust. Gas Exp.: 03/21/2027

Data Summary

Function	Result %BAC	Time hh:mm:ss	Barometric Pres. [hPa]	Comment(s) or Status Code
Ambient Air Blank	0.000	07:51:46		*TEST PASSED*
Control .04 Test 1 EC	0.040	07:52:21	1019	*TEST PASSED*
Control .04 Test 1 IR	0.040	07:52:21	1019	*TEST PASSED*
Ambient Air Blank	0.000	07:53:07		*TEST PASSED*
Control .04 Test 2 EC	0.041	07:53:19	1019	*TEST PASSED*
Control .04 Test 2 IR	0.040	07:53:19	1019	*TEST PASSED*
Ambient Air Blank	0.000	07:55:46		*TEST PASSED*
Control .08 Test 3 EC	0.080	07:56:21	1019	*TEST PASSED*
Control .08 Test 3 IR	0.080	07:56:21	1019	*TEST PASSED*
Ambient Air Blank	0.000	07:57:12		*TEST PASSED*
Control .08 Test 4 EC	0.081	07:57:24	1019	*TEST PASSED*
Control .08 Test 4 IR	0.081	07:57:24	1019	*TEST PASSED*
Ambient Air Blank	0.000	08:00:17		*TEST PASSED*
Control .16 Test 5 EC	0.160	08:00:55	1019	*TEST PASSED*
Control .16 Test 5 IR	0.161	08:00:55	1019	*TEST PASSED*
Ambient Air Blank	0.000	08:01:52		*TEST PASSED*
Control .16 Test 6 EC	0.162	08:02:07	1019	*TEST PASSED*
Control .16 Test 6 IR	0.163	08:02:07	1019	*TEST PASSED*
Ambient Air Blank	0.000	08:10:18		*TEST PASSED*
Control .30 Test 7 EC	0.298	08:10:54	1019	*TEST PASSED*
Control .30 Test 7 IR	0.300	08:10:54	1019	*TEST PASSED*
Ambient Air Blank	0.000	08:11:56		*TEST PASSED*
Control .30 Test 8 EC	0.301	08:12:09	1019	*TEST PASSED*
Control .30 Test 8 IR	0.304	08:12:09	1019	*TEST PASSED*
Ambient Air Blank	0.000	08:12:20		*TEST PASSED*

Result

All tests within acceptable tolerance.

Coordinator

Last Name: Bellay - First Name: David MI: M. Badge No.: 8112

On this date, I certified the above instrument in accordance with the Alcotest 9510 operator training and procedures established by the NJSP Office of Forensic Sciences.



Signed:

Date: 06/02/2026

ID: 50

ALCOTEST 9510 CYLINDER INSTALLATION REPORT - INLET 1
Toms River Twp
SERIAL NUMBER: ARMJ-0142

Equipment

Inst. Model No.: ALCOTEST 9510 Serial No.: ARMJ-0142
Firmware: 8326739 1.5 Config.: 8326737 3.10 WinCE: 8326738 2.9
Cyl1 Install File No.: 27 Cyl1 Install Date: 06/02/2026 Cyl1 Install No.: 1

Control Tests (0.100%)

Installation Inlet: #1 (Upper) Post test active Cyl.: #1 (Upper)
Dry Gas Lot No.: 302-403119242 Dry Gas Lot Exp.: 08/08/2027

Data Summary

Function	Result %BAC	Time hh:mm:ss	Barometric Pres. [hPa]	Comment(s) or Status Code
Ambient Air Blank	0.000	08:24:47		*TEST PASSED*
Control Test 1			1019	*TEST PASSED*
EC Result	0.101	08:25:33		*TEST PASSED*
IR Result	0.101	08:25:33		*TEST PASSED*
Ambient Air Blank	0.000	08:26:26		*TEST PASSED*
Control Test 2			1019	*TEST PASSED*
EC Result	0.102	08:26:50		*TEST PASSED*
IR Result	0.102	08:26:50		*TEST PASSED*
Ambient Air Blank	0.000	08:27:44		*TEST PASSED*
Control Test 3			1019	*TEST PASSED*
EC Result	0.102	08:28:08		*TEST PASSED*
IR Result	0.102	08:28:08		*TEST PASSED*
Ambient Air Blank	0.000	08:28:19		*TEST PASSED*

Result

All tests within acceptable tolerance.

Coordinator

Last Name: Bellay - First Name: David MI: M. Badge No.: 8112

On this date, I certified the above instrument in accordance with the Alcotest 9510 operator training and procedures established by the NJSP Office of Forensic Sciences.



Signed:

Date: 06/02/2026

ID: 50

ALCOTEST 9510 CYLINDER INSTALLATION REPORT - INLET 2
Toms River Twp
SERIAL NUMBER: ARMJ-0142

Equipment

Inst. Model No.:	ALCOTEST 9510	Serial No.:	ARMJ-0142		
Firmware:	8326739 1.5	Config.:	8326737 3.10	WinCE:	8326738 2.9
Cyl2 Install File No.:	28	Cyl2 Install Date:	06/02/2026	Cyl2 Install No.:	1

Control Tests (0.100%)

Installation Inlet:	#2 (Lower)	Post test active Cyl.:	#1 (Upper)
Dry Gas Lot No.:	302-403386733	Dry Gas Lot Exp.:	07/10/2028

Data Summary

Function	Result %BAC	Time hh:mm:ss	Barometric Pres. [hPa]	Comment(s) or Status Code
Ambient Air Blank	0.000	08:34:28		*TEST PASSED*
Control Test 1			1019	*TEST PASSED*
EC Result	0.102	08:35:13		*TEST PASSED*
IR Result	0.102	08:35:13		*TEST PASSED*
Ambient Air Blank	0.000	08:36:06		*TEST PASSED*
Control Test 2			1019	*TEST PASSED*
EC Result	0.102	08:36:30		*TEST PASSED*
IR Result	0.102	08:36:30		*TEST PASSED*
Ambient Air Blank	0.000	08:37:23		*TEST PASSED*
Control Test 3			1019	*TEST PASSED*
EC Result	0.103	08:37:47		*TEST PASSED*
IR Result	0.103	08:37:47		*TEST PASSED*
Ambient Air Blank	0.000	08:37:58		*TEST PASSED*

Result

All tests within acceptable tolerance.

Coordinator

Last Name: Bellay - First Name: David MI: M. Badge No.: 8112

On this date, I certified the above instrument in accordance with the Alcotest 9510 operator training and procedures established by the NJSP Office of Forensic Sciences.



Signed:

Date: 06/02/2026

ID: 50

CERTIFICATE OF ANALYSIS

EBS - ETHANOL BREATH STANDARD

Part Number: 4401036
DRAEGER MEDICAL SYSTEMS INC

Sales order: 1133220061
Date: September 10, 2024

METHOD OF ANALYSIS: IR Breath Alcohol Analyzer
ANALYTICAL ACCURACY: +/-0.002 BrAC or +/-2% whichever is greater.
CALGAZ LOT#: 302-403119242
ETHANOL IN NITROGEN

Manufactured Date: August 08, 2024
Product Expiration: August 08, 2027

COMPONENT	PPM	(BrAC)
ETHANOL	260.5PPM	(0.100)
NITROGEN	BAL	
AVERAGE ANALYTICAL VALUE	PPM	(BrAC)
ETHANOL	262.4	(0.101)

REFERENCE STANDARD	CYLINDER	CONCENTRATION PPM
N.M.I. TRACEABLE STANDARDS*	ND38424	260.7

* CERTIFICATION TRACEABLE TO NATIONAL METROLOGY INSTITUTE TRACEABLE STANDARDS

TRACEABILITY

Preparation:

Gas mixtures manufactured with balances calibrated by an ISO 17025 accredited company using NIST traceable weights and meets or exceeds the requirements of NIST Handbook 44.

Traceable certificate numbers 3445312 and 3398673.

Analytical:

Analytical Instruments Calibrated Using NMI Traceable Standards.


Certification Numbers: A679-20190918, D049803-20220329

No effecting environmental conditions during analysis.

*NMI is recognized by NIST through the Mutual Recognition Agreement (CIPM MRA).

CALGAZ calibration devices were found to meet all applicable requirements of the National Highway Traffic Safety Administration Model Specifications for calibrating units for breath alcohol testers.

APPROVED BY:



"We certify that all the cylinders for the Lot numbers identified herein are manufactured and tested within the requirements of CFR 49 part 178.65 and that physical and chemical test reports are on file and copies will be furnished upon request."

CALGAZ, a division of Airgas USA LLC
821 Chesapeake Drive, Cambridge, MD 21613-0149
Phone: (410) 228-6400 Fax: (410) 228-4251

CERTIFICATE OF ANALYSIS

EBS - ETHANOL BREATH STANDARD

Part Number: 4401036
DRAEGER MEDICAL SYSTEMS INC

Sales order: 1141350413
 Date: July 31, 2025

METHOD OF ANALYSIS: IR Breath Alcohol Analyzer
 ANALYTICAL ACCURACY: +/-0.002 BrAC or +/-2% whichever is greater.
 CALGAZ LOT#: 302-403386733
 ETHANOL IN NITROGEN

Manufactured Date: July 10, 2025
 Product Expiration: July 10, 2028

COMPONENT	PPM	(BrAC)
ETHANOL	260,5PPM	(0.100)
NITROGEN	BAL	

AVERAGE ANALYTICAL VALUE	PPM	(BrAC)
ETHANOL	261.3	(0.100)

REFERENCE STANDARD	CYLINDER	CONCENTRATION PPM
N.M.I. TRACEABLE STANDARDS*	ND49826	260.1

* CERTIFICATION TRACEABLE TO NATIONAL METROLOGY INSTITUTE TRACEABLE STANDARDS

TRACEABILITY

Preparation:

Gas mixtures manufactured with balances calibrated by an ISO 17025 accredited company using NIST traceable weights and meets or exceeds the requirements of NIST Handbook 44. Traceable certificate numbers 3445312 and 3398673.


Analytical:

Analytical Instruments Calibrated Using NMI Traceable Standards.
 Certification Numbers: A679-20190918, D049803-20220329

No effecting environmental conditions during analysis.

*NMI is recognized by NIST through the Mutual Recognition Agreement (CIPM MRA).

CALGAZ calibration devices were found to meet all applicable requirements of the National Highway Traffic Safety Administration Model Specifications for calibrating units for breath alcohol testers.

APPROVED BY: 

"We certify that all the cylinders for the Lot numbers identified herein are manufactured and tested within the requirements of CFR 49 part 178.65 and that physical and chemical test reports are on file and copies will be furnished upon request."

CALGAZ, a division of Airgas USA LLC
 821 Chesapeake Drive, Cambridge, MD 21613-0149
 Phone: (410) 228-6400 Fax: (410) 228-4251



Dräger

Alcotest 9510

CERTIFICATE OF ACCURACY

This is to certify that the Alcotest 9510 has been tested for accuracy and found to be in compliance with the National Highway Traffic Safety Administration Standard for evidential breath testing devices. The Alcotest 9510 is compliant as a "mobile" and "nonmobile" EBT with 49 FR 48854, 49 FR 48864, and 58 FR 48705. The manufacturer recommends accuracy verification of this instrument within 12 months of the calibration date below, or sooner, according to your state's specifications.

Certification Date: Serial Number:

4/7/2026 ARMJ-0142

Draeger, Inc. *LMH*



State of New Jersey

OFFICE OF THE ATTORNEY GENERAL
DEPARTMENT OF LAW AND PUBLIC SAFETY
DIVISION OF STATE POLICE
POST OFFICE BOX 7068
WEST TRENTON, NJ 08628-0068
(609) 882-2000

PHILIP D. MURPHY
Governor

TAHESHA L. WAY
Lt. Governor

MATTHEW J. PLATKIN
Attorney General

COLONEL PATRICK J. CALLAHAN
Superintendent

CERTIFICATION OF ANALYSIS
0.100 PERCENT BREATH ALCOHOL SIMULATOR SOLUTION

ACCEPTANCE SPECIFICATIONS FOR BREATH ALCOHOL SIMULATOR SOLUTION: Ethyl alcohol concentration within, but not exceeding, the range of 0.1174 to 0.1246 grams per 100 milliliters of solution.

MANUFACTURER: Draeger, Inc.

ANALYSIS DATE: 07/01/2025

BREATH ALCOHOL SIMULATOR SOLUTION LOT NUMBER: 25190

Representative samples of the above-referenced Lot Number were tested by Gas Chromatography and found to have a mean ethyl alcohol concentration range of 0.1208 to 0.1219 grams per 100 milliliters of solution.

This lot of breath alcohol simulator solution may be utilized as a known traceable standard for the purpose of conducting periodic tests, pursuant to N.J.A.C. 13:51-4.3, of approved breath test instruments (N.J.A.C. 13:51-3.5) utilized by law enforcement agencies in this State. The manufacturer's expiration date for this lot of breath alcohol simulator solution is June 10, 2027.

As OFS Director for the Division of State Police, I hereby certify and attest that the tests and results documented in this Certificate of Analysis were performed at the Office of Forensic Sciences of the Division of State Police on properly functioning and calibrated instruments and equipment. All procedures utilized are accurate, objective, and performed on a routine basis by personnel within the Office of Forensic Sciences, in accordance with their professional duties and responsibilities.

Michael Kennedy
Michael Kennedy
Director
NJSP Office of Forensic Sciences

Sworn to and subscribed before me this 8 day of July, 2025.
Notary

KAREN E. STAHL
Notary Public, State of New Jersey
Comm. # 50110522
My Commission Expires 8/13/2029



An Internationally Accredited Agency
New Jersey Is An Equal Opportunity Employer
Printed on Recycled Paper and Recyclable



CALIBRATED BY TRANSCAL CERTIFICATE OF CALIBRATION

Customer: DRAEGER INC
 7256 S SAM HOUSTON PKWY W
 STE 100
 HOUSTON, TX 77085
PO Number: 4303848314

Customer: DRAEGER INC
 7256 S SAM HOUSTON PKWY W
 STE 100
 HOUSTON, TX 77085
PO Number: 4303848314



Certificate/SO Number: 5-G0P3R-40-1 Revision 0

Manufacturer: Drager Safety AG & Co. KGaA
Model Number: X-Cal 2000
Description: Breath Alcohol Simulator
Serial Number: ARRL-0020
ID: NONE

As-Found: In Tolerance
As-Left: In Tolerance

Issue Date: Jan 15, 2026
Calibration Date: Jan 15, 2026
Due Date: Jan 15, 2027

Calibrated To: Manufacturer Specification
Calibration Procedure: 1-AC103519-2

This calibration conforms to ISO/IEC 17025:2017, ANSI/NCSL Z540-1-1994, ISO 9001:2015, and ISO 10012:2003. When specified contractually, the requirements of ISO TSI6949:2009, 10CFR21, 10CFR50 App. B, ASME NQA-1:2012, and ANSI/NCSL Z540.3-2008 (R2013) are also covered. Standards are traceable to the SI either through the National Institute of Standards & Technology (NIST), the National Research Council of Canada (NRC), other national measurement institutes (NMI) that are signatories to the CIPM Mutual Recognition Arrangement, natural physical constants, intrinsic standards, or ratio type measurements. The results of this report relate only to the item calibrated or tested. Reported data is valid at the time of calibration at the environmental conditions noted. This certificate shall not be reproduced except in full, without the written consent of Transcal Inc. Uncertainties are reported with a coverage factor k=2, providing a level of confidence of approximately 95%. When a decision rule is inherent in the specification or the standard the prescribed decision rule was used, otherwise where a statement of conformance is made, it was made based on the measurement meeting the limit, with no guard-banding applied. Per ILAC G8 Simple Decision rule TUR is $k=4:1$ unless otherwise noted in the data.

Calibrated At:
 16115 Park Row
 Houston, TX 77084

Facility Responsible:
 16115 Park Row
 Houston, TX 77084
 800-828-1470

Calibrated By:
 Jempson Andre
 Houston, TX 77084

Facility Responsible:
 16115 Park Row
 Houston, TX 77084
 800-828-1470

Calibrated By:
 Electronically Signed By:
 Jempson Andre

Reviewed By:
 Electronically Signed By:
 Graham Walker for

Unit Barcode: 
 09908665740

Jempson Andre	Jan 15, 2026	Josh Soileau	Jan 15, 2026
Calibration Technician	03:36:31 -05:00	Lab Manager	06:13:44 -05:00

Date Received: January 14, 2026
Service Level: R9

Certificate - Page 1 of 2

Customer Number: 1-659111-00C

Customer: DRAEGER INC
7256 S SAM HOUSTON PKWY W
STE 100
HOUSTON, TX 77085
PO Number: 4303848314



Certificate/SO Number: 5-G0P3R-40-1 Revision 0

As Found/As Left Data

Description	Setpoints	Accuracy	Low Limit	High Limit	As Found / As Left	O U T	Cal Process Uncertainty (k=2; ±)	Measurement Uncertainty (k=2; ±)	Units	TUR
Function Checks										
Bubble Check			P	P	P					
Seal Check			P	P	P					
Temperature Source: Accuracy Test										
Accuracy Test	34.00°C	±(0.02 °C)	33.98	34.02	34.00 °C		1.5e-002	1.6e-002	°C	1.3:1
Temperature Source: Stability Test										
Stability Test	0.00°C	±(0.02°C)	0.02	0.02	0.00 °C		1.5e-002	1.6e-002	°C	1.3:1

Field not applicable.

Traceable Standards

Asset	Manufacturer	Model Number	Description	Cal Date	Due Date	Traceability Number	Use
05H1428	AccuMac Corporation	AM1760-12-P	Secondary SPRT	18-Apr-25	30-Apr-26	15-805H1428-4-1	AF/AL
5656	Fluke	1524	Handheld Thermometer	13-Jun-25	30-Jun-26	8-85656-21-1	AF/AL

The use of the standard is defined as: AF - used for as-found readings, AL - used for as-left readings.


Environmental Data

Temperature	Relative Humidity	Temp / RH Asset	Lab Area	Lab Description
71.07°F /21.71°C	29.70%	DewK11	G	Temperature

CALIBRATED
BY TRANSCAT

CERTIFICATE OF CALIBRATION

Customer: DRAEGER INC
7256 S SAM HOUSTON PKWY W
STE 100
HOUSTON, TX 77085
PO Number: SUC4303700862



Certificate/SO Number: 5-F8B2G-40-1 Revision 0

Manufacturer: Mensor Corp
Model Number: CP02500
Description: Portable Barometer
Serial Number: 41001275
ID: NONE

As-Found: Out Of Tolerance
As-Left: In Tolerance

Issue Date: Aug 25, 2025
Calibration Date: Aug 23, 2025
Due Date: Aug 23, 2026

Calibrated To: Manufacturer Specification
Calibration Procedure: 1-AC107288-0

Transcat Calibration Laboratories have been audited and found in compliance with ISO/IEC 17025:2017. Accredited calibrations performed within the Lab Scope of Accreditation are indicated by the presence of the Accrediting Body Logo and Certificate Number. Any measurements on an accredited calibration not covered by the Lab Scope of Accreditation are listed in the notes section of the certificate. SDG, NRC, CLAS or ANAB do not guarantee the accuracy of an individual calibration by accredited laboratories.

Transcat calibrations, as applicable, are performed in compliance with the requirements of the Transcat Quality Manual QAC-P01-000, the customer Purchase Order and/or Quality Agreement requirements, ISO 9001:2015, ANSI/NISO Z540.1-1994 (R2002), and ISO 10012:2003, as applicable. When specified contractually, the requirements of ISO TS 16949:2009, 16CFR21, 16CFR60 App. 6, AECME NQA-1:2012, and ANSI/NCCL Z540.3-2006 (R2012) are also covered.

Complete records of work performed are maintained by Transcat and are available for inspection. Laboratory standards used in the performance of this calibration are listed on this certificate.

Transcat documents the traceability of measurements to the SI units through the National Institute of Standards and Technology (NIST), or the National Research Council of Canada (NRC), or other national measurement institutes (NMI) that are signatories to the OIML Mutual Recognition Arrangement, or accepted fundamental and/or natural physical constants, or by the use of specified methods, consensus standards or ratio type measurements. Documentation supporting traceability information is available for review upon written request at a Transcat facility. The measured quantity and the measurement uncertainty are required for further dissemination of traceability.

Uncertainties are reported with a coverage factor $k=2$, providing a level of confidence of approximately 95%. All calibrations have been performed using processes having a TUR of 4:1 or better (3:1 for mass calibrations), unless otherwise noted. The Test Uncertainty Ratio (TUR) is calculated in accordance with NCSL International RP-19. For mass calibrations: Conventional mass referenced to 8.0 g/cm³.

The results in this report relate only to the item calibrated or tested. Recorded calibration data is valid at the time of calibration within the stated uncertainties at the environmental conditions noted. The determination of compliance to the specification is specific to the manufacturer's ID on, referenced above based on the tolerances shown; these tolerances are either the original equipment manufacturer's (OEM's) warranted specifications or the client's requested specifications. Any number of factors can cause a unit to drift out of tolerance at any time following its calibration. Limitations on the use of this instrument are detailed in the OEM's operating instructions. This certificate may not be reproduced except in full, without the written approval of Transcat. Additional information, if applicable may be included on separate report(s).

Notes:

Unit received out of tolerance. Adjustments were made to meet customer specs. The COI readings were verified.

Date Received: August 08, 2025
Service Level: R3


Certificate - Page 1 of 6
Reprinted on August 27, 2025

Customer Number: 1-859-111-000
OPS-F20-014R11 02/27/23 FP001R9 4/3/2024

CALIBRATED
BY **TRANSOT**

CERTIFICATE OF CALIBRATION

Customer: DRAEGER INC
7256 S SAM HOUSTON PKWY W
STE 100
HOUSTON, TX 77085
PO Number: SUC4303700852



Certificate/SO Number: 5-F8B2G-40-1 Revision 0

As Found Data

Description	Setpoints	Accuracy	Low Limit	High Limit	As Found	O O T	Cal Process Uncertainty (k=2; s)	Measurement Uncertainty (k=2; s)	Units	TUR
Pressure Measure: 552 to 1172 mbars Range										
	550.07mbar	±(0.015% FS)	549.8	550.3	549.7 mbar	*	1.0e-002	5.1e-002	mbar	19.1 : 1
	610.01mbar	±(0.015% FS)	599.8	610.2	609.7 mbar	*	1.2e-002	5.9e-002	mbar	17.9 : 1
	680.38mbar	±(0.015% FS)	680.2	680.6	680.0 mbar	*	1.3e-002	5.9e-002	mbar	15.8 : 1
	734.28mbar	±(0.015% FS)	734.1	734.5	733.9 mbar	*	1.4e-002	5.9e-002	mbar	14.3 : 1
	804.64mbar	±(0.015% FS)	804.4	804.8	804.3 mbar	*	1.6e-002	6.0e-002	mbar	13.1 : 1
	864.92mbar	±(0.015% FS)	864.7	865.1	864.6 mbar	*	1.6e-002	6.0e-002	mbar	12.2 : 1
	924.82mbar	±(0.015% FS)	924.7	925.1	924.6 mbar	*	1.6e-002	6.0e-002	mbar	11.4 : 1
	986.22mbar	±(0.015% FS)	986.0	985.4	984.9 mbar	*	1.9e-002	6.1e-002	mbar	10.7 : 1
	1043.85mbar	±(0.015% FS)	1043.7	1044.1	1043.5 mbar	*	2.0e-002	6.1e-002	mbar	10.1 : 1
	1114.20mbar	±(0.015% FS)	1114.0	1114.4	1113.8 mbar	*	2.1e-002	6.1e-002	mbar	9.4 : 1
	1174.68mbar	±(0.015% FS)	1174.4	1174.8	1174.2 mbar	*	2.2e-002	6.2e-002	mbar	9.0 : 1
	924.92mbar	±(0.015% FS)	924.7	925.1	924.6 mbar	*	1.6e-002	6.0e-002	mbar	11.4 : 1
	864.92mbar	±(0.015% FS)	864.7	865.1	864.6 mbar	*	1.6e-002	6.0e-002	mbar	12.2 : 1
	804.64mbar	±(0.015% FS)	804.4	804.8	804.3 mbar	*	1.6e-002	6.0e-002	mbar	13.1 : 1

Date Received: August 08, 2025
Service Level: RS


Certificate - Page 2 of 5
Reprinted on August 27, 2025

Customer Number: 1-859111-000
QPS-F20-014(1) 07/27/23 FF001R9 4/9/2021

CALIBRATED
BY TRANSCOT

CERTIFICATE OF CALIBRATION

Customer: DRAEGER INC
7256 S SAM HOUSTON PKWY W
STE 100
HOUSTON, TX 77085
PO Number: SUC4303700862



Certificate/SO Number: 5-F8B2G-40-1 Revision 0

As Left Data

Description	Setpoints	Accuracy	Low Limit	High Limit	As Left	O C T	Cal Process Uncertainty (k=2; ±)	Measurement Uncertainty (k=2; ±)	Units	TUR
Pressure Measure: 552 to 1172 mbara Range										
	550.07mbara	±(0.015% FS)	548.9	550.3	550.1 mbara		1.0e-002	6.1e-002	mbara	19.1:1
	619.01mbara	±(0.015% FS)	609.8	610.2	619.1 mbara		1.2e-002	5.6e-002	mbara	17.9:1
	680.37mbara	±(0.015% FS)	680.2	680.5	680.4 mbara		1.3e-002	5.9e-002	mbara	15.5:1
	734.27mbara	±(0.015% FS)	734.1	734.5	734.3 mbara		1.4e-002	5.3e-002	mbara	14.3:1
	804.64mbara	±(0.015% FS)	804.4	804.8	804.7 mbara		1.5e-002	6.0e-002	mbara	13.1:1
	864.91mbara	±(0.015% FS)	864.7	865.1	865.0 mbara		1.6e-002	6.0e-002	mbara	12.2:1
	924.91mbara	±(0.015% FS)	924.7	925.1	925.0 mbara		1.8e-002	6.0e-002	mbara	11.4:1
	985.21mbara	±(0.015% FS)	985.0	985.4	985.3 mbara		1.9e-002	6.1e-002	mbara	10.7:1
	1043.84mbara	±(0.015% FS)	1043.6	1044.0	1043.9 mbara		2.0e-002	6.1e-002	mbara	10.1:1
	1114.19mbara	±(0.015% FS)	1114.0	1114.4	1114.2 mbara		2.1e-002	6.1e-002	mbara	9.4:1
	1174.55mbara	±(0.015% FS)	1174.4	1174.8	1174.8 mbara		2.2e-002	6.2e-002	mbara	9.0:1
	824.91mbara	±(0.015% FS)	824.7	825.1	825.0 mbara		1.8e-002	6.0e-002	mbara	11.4:1
	864.91mbara	±(0.015% FS)	864.7	865.1	865.0 mbara		1.8e-002	6.0e-002	mbara	12.2:1
	804.63mbara	±(0.015% FS)	804.4	804.8	804.7 mbara		1.5e-002	6.0e-002	mbara	15.1:1

Field not applicable.

Traceable Standards

Asset	Manufacturer	Model Number	Description	Cal Date	Due Date	Traceability Number	Use
RW11BA	Fluke/DH Instruments	PG37601	Piston Gauge	31-Jul-25	31-Jul-26	5-3DW11BA-2b-1	AF/AL
DW11CA	Fluke/DH Instruments	MS-AMH-38	AMH Mass Set	5-Jun-25	30-Sep-25	5-3DW11CA-4b-1	AF/AL
DW11LOW	Fluke/DH Instruments	PC-7100/7600-1b-TC	Gas Piston-Cylinder Module	8-Apr-22	30-Apr-27	5-3DW11LOW-3-1	AF/AL
DW11MASS	Fluke/DH Instruments	MS-AMH-38	AMH Mass Set	5-Mar-25	31-Mar-26	5-3DW11MASS-12-1	AF/AL

The use of the standard is defined as: AF - used for as-found readings, AL - used for as-left readings.

Date Received: August 05, 2025
Service Level: R9


Certificate - Page 3 of 6
Reprinted on August 27, 2025

Customer Number: 1-689111-000
OPS-F20-014311 07/27/23 FPO01R6 43/2021

CALIBRATED
BY TRANSCAT

CERTIFICATE OF CALIBRATION

Customer: DRAEGER INC
7265 S SAM HOUSTON PKWY W
STE 100
HOUSTON, TX 77085
PO Number: SUC4309700862



Certificate/SO Number: 5-F8B2G-40-1 Revision 0

Environmental Data

Temperature	Relative Humidity	Temp / RH Asset	Lab Area	Lab Description
73.10°F / 22.83°C	43.00%	DewK9	B	GP Pressure

Decision Rule

When compliance statements are present, they are reported without factoring in the effects of uncertainty and comply with the guidelines as follows: The acceptance zone is defined as: less than or equal to the high limit, and/or greater than or equal to the low limit. The rejection zones are defined as greater than the high limit and/or less than the low limit. Single measurement results in the acceptance zone are identified as in-tolerance. Single measurement results in the rejection zone are identified as out-of-tolerance (OOT). When all measurement results are in the acceptance zone for repeated measurements, for the same characteristic, the test is identified as in-tolerance. For repeated characteristic measurements, a single measurement result in the rejection zone, will cause the test to be identified as out-of-tolerance (OOT). Data rejection for cause, (outliers) is permitted after the "Determining and Verifying Out Of Tolerance (OOT) and/or Op Fail Readings" procedure outlined in this document has been completed and the anomalous reading cannot be repeated, and the anomalous reading does not represent the system under test. Statements of conformity are binary.


Date Received: August 08, 2023
Service Level: R9

Certificate - Page 4 of 6
Reprinted on August 27, 2025

Customer Number: 1-859-111-000
DPS-F20-014R11 07/27/23 FP001R0 4/3/2021

CALIBRATED BY TRANSCOT CERTIFICATE OF CALIBRATION

Customer: DRAEGER INC
 7256 S SAM HOUSTON PKWY W
 STE 100
 HOUSTON, TX 77085
 PO Number: SUC4303700862



Certificate/SO Number: 5-F8B2G-40-1 Revision 0

Legend

Topic	Description
Accuracy	UUT specification that establishes expected tolerances and a time limit (calibration interval) over which the instrument is expected to hold these tolerances
As Found	Initial measurement results
As Left	Measurement results after adjustment and/or repair
Blank Data Field	Test is not applicable for the UUT
Cal Process Uncertainty (CPU)	The uncertainty of calibration process for the reported measurement result
Calibration Date	Indicates the date that the calibration was completed
Cover Factor (k)	A measure of uncertainty that defines an interval about the measurement result
Due Date	Indicates the end of the calibration cycle as requested by the customer
Issue Date	Indicates the date that the calibration has passed the Data Review Process and was signed by an authorized signatory or the date that a revision to the original certificate has been issued
Low/High Limits	Establishes UUT acceptable performance limits for the test measurement
Measurement Uncertainty	The dispersion of the values attributed to a measured quantity
OOA	Out of Acceptance (#)
OOT	Out of Tolerance (*)
Setpoints	Measurement target values
Traceability	Unbroken chain of comparisons relating an instrument's measurements to a known standard(s)
Traceability Number	Unique identifier(s) used to document traceability of calibration standard(s)
TUR	Test Uncertainty Ratio, ratio of the tolerance or specification of the test measurement in relation to the uncertainty in measurement results
UUT	Unit Under test

Date Received: August 08, 2025
 Service Level: R3

Certificate - Page 5 of 6
 Reprinted on August 27, 2025

Customer Number: 1-659111-000
 OPS-F20-014R11 07/27/23 FP001R0 4/9/2021

CALIBRATED
BY **TRANSCAT**

CERTIFICATE OF CALIBRATION


Customer: DRAEGER INC
7256 S SAM HOUSTON PKWY W
STE 100
HOUSTON, TX 77085
PO Number: SUC4303700862



Certificate/SO Number: 5-F8B2G-40-1 Revision 0


Calibrated At:
16115 Park Row
Houston, TX 77084

Facility Responsible:
16115 Park Row
Houston, TX 77084
800-828-1470


Unit Barcode: 
09008426643

Date Received: August 08, 2025
Service Level: R0

Certificate - Page 6 of 6
Reprinted on August 27, 2025

Calibrated By:
 Electronically Signed By:
Alex Spilker

Alex Spilker Aug 23, 2025
Calibration Technician 20:24:59 -04:00

Reviewed By:
 Electronically Signed By:
Graham Walker for

Josh Sellen Aug 25, 2025
Lab Manager 08:31:52 -04:00

Customer Number: 1-850-111-000
OPS-P20-014R11 07/27/23 FP001R04/09/2021

CERTIFICATE OF ANALYSIS

EBS - ETHANOL BREATH STANDARD

Part Number: 4401036
DRAEGER MEDICAL SYSTEMS INC

Sales order: 1130434779
Date: May 23, 2024

METHOD OF ANALYSIS: IR Breath Alcohol Analyzer

ANALYTICAL ACCURACY: +/-0.002 BrAC or +/-2% whichever is greater.

CALGAZ LOT#: 302-403034217

Manufactured Date: April 30, 2024

ETHANOL IN NITROGEN

Product Expiration: April 30, 2027

COMPONENT	PPM	(BrAC)
ETHANOL	260.5PPM	(0.100)
NITROGEN	BAL	
AVERAGE ANALYTICAL VALUE	PPM	(BrAC)
ETHANOL	262.3	(0.101)

REFERENCE STANDARD	CYLINDER	CONCENTRATION PPM
N.M.I. TRACEABLE STANDARDS*	ND38424	260.7

* CERTIFICATION TRACEABLE TO NATIONAL METROLOGY INSTITUTE TRACEABLE STANDARDS

TRACEABILITY

Preparation:

Gas mixtures manufactured with balances calibrated by an ISO 17025 accredited company using NIST traceable weights and meets or exceeds the requirements of NIST Handbook 44.

Traceable certificate numbers 3445312 and 3398673.

Analytical:

Analytical Instruments Calibrated Using NMI Traceable Standards.


Certification Numbers: A679-20190918, D049803-20220329

No effecting environmental conditions during analysis.

*NMI is recognized by NIST through the Mutual Recognition Agreement (CIPM MRA).

CALGAZ calibration devices were found to meet all applicable requirements of the National Highway Traffic Safety Administration Model Specifications for calibrating units for breath alcohol testers.

APPROVED BY:



"We certify that all the cylinders for the Lot numbers identified herein are manufactured and tested within the requirements of CFR 49 part 178.65 and that physical and chemical test reports are on file and copies will be furnished upon request."

CALGAZ, a division of Airgas USA LLC
821 Chesapeake Drive, Cambridge, MD 21613-0149
Phone: (410) 228-6400 Fax: (410) 228-4251

CERTIFICATE OF ANALYSIS

EBS - ETHANOL BREATH STANDARD

Part Number: 4507062
DRAEGER MEDICAL SYSTEMS INC

Sales order: 1130434779
Date: May 22, 2024

METHOD OF ANALYSIS: IR-Breath-Alcohol-Analyzer
ANALYTICAL ACCURACY: +/-0.002 BrAC or +/-2% whichever is greater.
CALGAZ LOT#: 302-402999655
ETHANOL IN NITROGEN

Manufactured Date: March 20, 2024
Product Expiration: March 20, 2027

COMPONENT	PPM	(BrAC)
ETHANOL	104.2PPM	(0.040)
NITROGEN	BAL	
AVERAGE ANALYTICAL VALUE	PPM	(BrAC)
ETHANOL	107.9	(0.041)

REFERENCE STANDARD	CYLINDER	CONCENTRATION PPM
N.M.I. TRACEABLE STANDARDS*	ND28529	103.7

N.M.I. TRACEABLE STANDARDS*

ND28529

103.7

* CERTIFICATION TRACEABLE TO NATIONAL METROLOGY INSTITUTE TRACEABLE STANDARDS

TRACEABILITY

Preparation:

Gas mixtures manufactured with balances calibrated by an ISO 17025 accredited company using NIST traceable weights and meets or exceeds the requirements of NIST Handbook 44.

Traceable certificate numbers 3445312 and 3398673.

Analytical:

Analytical Instruments Calibrated Using NMI Traceable Standards.


Certification Numbers: A679-20190918, D049803-20220329

No effecting environmental conditions during analysis.

*NMI is recognized by NIST through the Mutual Recognition Agreement (CIPM MRA).

CALGAZ calibration devices were found to meet all applicable requirements of the National Highway Traffic Safety Administration Model Specifications for calibrating units for breath alcohol testers.

APPROVED BY:



"We certify that all the cylinders for the Lot numbers identified herein are manufactured and tested within the requirements of CFR 49 part 178.55 and that physical and chemical test reports are on file and copies will be furnished upon request."

CALGAZ, a division of Airgas USA LLC

821 Chesapeake Drive, Cambridge, MD 21613-0149

Phone: (410) 228-8400 Fax: (410) 228-4251

CERTIFICATE OF ANALYSIS

EBS - ETHANOL BREATH STANDARD

Part Number: 4401032
DRAEGER MEDICAL SYSTEMS INC

Sales order: 1130434779
Date: May 22, 2024

METHOD OF ANALYSIS: IR Breath Alcohol Analyzer

ANALYTICAL ACCURACY: +/-0.002 BrAC or +/-2% whichever is greater.

CALGAZ LOT#: 302-403008098

Manufactured Date: April 01, 2024

ETHANOL IN NITROGEN

Product Expiration: April 01, 2027

COMPONENT	PPM	(BrAC)
ETHANOL	208.4PPM	(0.080)
NITROGEN	BAL	
AVERAGE ANALYTICAL VALUE	PPM	(BrAC)
ETHANOL	212.7	(0.082)

REFERENCE STANDARD	CYLINDER	CONCENTRATION PPM
N.M.I. TRACEABLE STANDARDS*	ND28529	103.7

* CERTIFICATION TRACEABLE TO NATIONAL METROLOGY INSTITUTE TRACEABLE STANDARDS

TRACEABILITY

Preparation:

Gas mixtures manufactured with balances calibrated by an ISO 17025 accredited company using NIST traceable weights and meets or exceeds the requirements of NIST Handbook 44.

Traceable certificate numbers 3445312 and 3398673.

Analytical:

Analytical Instruments Calibrated Using NMI Traceable Standards.


Certification Numbers: A679-20190918, D049803-20220329

No effecting environmental conditions during analysis.

**NMI is recognized by NIST through the Mutual Recognition Agreement (CIPM MRA).*

CALGAZ calibration devices were found to meet all applicable requirements of the National Highway Traffic Safety Administration Model Specifications for calibrating units for breath alcohol testers.

APPROVED BY:



"We certify that all the cylinders for the Lot numbers identified herein are manufactured and tested within the requirements of CFR 49 part 178.65 and that physical and chemical test reports are on file and copies will be furnished upon request."

CALGAZ, a division of Airgas USA LLC

821 Chesapeake Drive, Cambridge, MD 21613-0149

Phone: (410) 228-6400

Fax: (410) 228-4251

CERTIFICATE OF ANALYSIS

EBS - ETHANOL BREATH STANDARD

Part Number: 4401040NJ
DRAEGER MEDICAL SYSTEMS INC

Sales order: 1130434779
Date: May 22, 2024

METHOD OF ANALYSIS: IR Breath Alcohol Analyzer
ANALYTICAL ACCURACY: +/-0.002 BrAC or +/-2% whichever is greater.
CALGAZ LOT#: 302-402997176
ETHANOL IN NITROGEN

Manufactured Date: March 19, 2024
Product Expiration: March 19, 2027

COMPONENT	PPM	(BrAC)
ETHANOL	416.8PPM	(0.160)
NITROGEN	BAL	
AVERAGE ANALYTICAL VALUE	PPM	(BrAC)
ETHANOL	423.2	(0.162)

REFERENCE STANDARD	CYLINDER	CONCENTRATION PPM
N.M.I. TRACEABLE STANDARDS*	ND23368	519.5

* CERTIFICATION TRACEABLE TO NATIONAL METROLOGY INSTITUTE TRACEABLE STANDARDS

TRACEABILITY

Preparation:

Gas mixtures manufactured with balances calibrated by an ISO 17025 accredited company using NIST traceable weights and meets or exceeds the requirements of NIST Handbook 44.

Traceable certificate numbers 3445312 and 3398673.

Analytical:

Analytical Instruments Calibrated Using NMI Traceable Standards.


Certification Numbers: A679-20190918, D049803-20220329

No effecting environmental conditions during analysis.

*NMI is recognized by NIST through the Mutual Recognition Agreement (CIPM MRA).

CALGAZ calibration devices were found to meet all applicable requirements of the National Highway Traffic Safety Administration Model Specifications for calibrating units for breath alcohol testers.

APPROVED BY: _____



"We certify that all the cylinders for the Lot numbers identified herein are manufactured and tested within the requirements of CFR 49 part 178.05 and that physical and chemical test reports are on file and copies will be furnished upon request."

CALGAZ, a division of Airgas USA LLC
821 Chesapeake Drive, Cambridge, MD 21613-0149
Phone: (410) 228-6400 Fax: (410) 228-4251

CERTIFICATE OF ANALYSIS

EBS - ETHANOL BREATH STANDARD

Part Number: 4401041NJ

Sales order: 1130434779

DRAEGER MEDICAL SYSTEMS INC

Date: May 22, 2024

METHOD OF ANALYSIS: IR Breath Alcohol Analyzer

ANALYTICAL ACCURACY: +/-0.002 BrAC or +/-2% whichever is greater.

Manufactured Date: March 21, 2024

CALGAZ LOT#: 302-403001320

ETHANOL IN NITROGEN

Product Expiration: March 21, 2027

COMPONENT	PPM	(BrAC)
ETHANOL	781.5PPM	(0.300)
NITROGEN	BAL	
AVERAGE ANALYTICAL VALUE	PPM	(BrAC)
ETHANOL	784.1	(0.301)

REFERENCE STANDARD	CYLINDER	CONCENTRATION PPM
N.M.I. TRACEABLE STANDARDS*	ND23368	519.5

* CERTIFICATION TRACEABLE TO NATIONAL METROLOGY INSTITUTE TRACEABLE STANDARDS

TRACEABILITY

Preparation:

Gas mixtures manufactured with balances calibrated by an ISO 17025 accredited company using NIST traceable weights and meets or exceeds the requirements of NIST Handbook 44.

Traceable certificate numbers 3445312 and 3398673.

Analytical:


Analytical Instruments Calibrated Using NMI Traceable Standards.

Certification Numbers: A679-20190918, D049803-20220329

No effecting environmental conditions during analysis.

*NMI is recognized by NIST through the Mutual Recognition Agreement (CIPM MRA).

CALGAZ calibration devices were found to meet all applicable requirements of the National Highway Traffic Safety Administration Model Specifications for calibrating units for breath alcohol testers.

APPROVED BY: 

"We certify that all the cylinders for the Lot numbers identified herein are manufactured and tested within the requirements of CFR 49 part 178.65 and that physical and chemical test reports are on file and copies will be furnished upon request."

CALGAZ, a division of Airgas USA LLC
 821 Chesapeake Drive, Cambridge, MD 21613-0149
 Phone: (410) 228-6400 Fax: (410) 228-4251

DEPARTMENT OF
Traffic and Public Safety
This is to certify that

David M. Bellay

New Jersey State Police

IS QUALIFIED AND COMPETENT TO CONDUCT CHEMICAL BREATH ANALYSIS PURSUANT TO CHAPTER 142 OF

THE LAWS OF 1964 IN THE OPERATION OF THE Alcotest 9510

A METHOD TO DETERMINE INTOXICATION

GIVEN UNDER MY HAND AT TRENTON, NEW JERSEY THIS 28th DAY OF April

TWO THOUSAND AND Twenty Three

David M. Bellay
COLONEL
NEW JERSEY STATE POLICE

Mr. J. A.
ATTORNEY GENERAL
STATE OF NEW JERSEY

ORIGINAL COURSE DATES

	DATE	Refresher Course PLACE	INSTRUCTOR
1.	3-27-25	MCFA	A. J.
2.			
3.			
4.			
5.			
6.			
7.			
8.			
9.			

S.P. 2938 (Rev. 10/22)

DEPARTMENT OF
Traffic and Public Safety
This is to certify that

David M. Bellay

Breath Test Coordinator/Instructor

IS QUALIFIED AND COMPETENT TO CONDUCT CHEMICAL BREATH ANALYSIS PURSUANT TO CHAPTER 142 OF

THE LAWS OF 1964 IN THE OPERATION OF THE Alcotest 9510

A METHOD TO DETERMINE INTOXICATION

GIVEN UNDER MY HAND AT TRENTON, NEW JERSEY THIS 26th DAY OF August

TWO THOUSAND AND Twenty Four

David M. Bellay
COLONEL
NEW JERSEY STATE POLICE

Mr. J. A.
ATTORNEY GENERAL
STATE OF NEW JERSEY

ORIGINAL COURSE DATES

	DATE	Refresher Course PLACE	INSTRUCTOR
1.			
2.			
3.			
4.			
5.			
6.			
7.			
8.			
9.			

S.P. 2938 (Rev. 10/22)